



GLOSS-GLOSSMETER

DETERMINATION OF GLOSS VALUE AT 20°, 60° AND 85°

The standard specifies a method for determining the gloss of coatings using the three geometries of 20°, 60° or 85°. The method is suitable for the gloss measurement of non-textured coatings on plane, opaque substrates.

